

# PERSONAL SERVICE SETTINGS PERFORMANCE DISCLOSURE PROGRAM



- Provide clients with verbal and written after-care instructions for invasive procedures
- Maintain complete client records for invasive procedures

- Operate and maintain the sterilizer and ultrasonic cleaner properly
- Check the sterilizer every 14 days with a biological spore test
- Report failed spore tests to Peel Public Health at 905-799-7700, ext. 6300
- Maintain complete sterilization records

- Use ear lobe piercing devices, micropigmentation equipment, sterilizers and other equipment used for invasive services that are permitted for use by Peel Public Health

- Provide cleaning and disinfecting supplies
- Clean with detergent and water using a scrub brush, then disinfect by fully immersing equipment in an approved disinfectant for the required contact time
- Know your disinfectants. Use an appropriate level of disinfectant for your equipment. Ask your Public Health Inspector for assistance.
- Follow manufacturer's instructions for disinfectant use, dilution, contact time, expiry date and disposal

- Wash your hands before and after providing services
- Wear gloves for invasive procedures
- Provide a sink with hot and cold running water, soap and single-use towels that is accessible at all times

- Provide and use prepackaged sterile equipment and jewellery
- Provide a sharps container to safely discard used needles, razors and lancets immediately after use

- Provide and use new single-use equipment for every client
- Discard used single-use equipment immediately after use on a client

- Do not double dip from the original container
- Use a new single-use applicator for every dip

- Clean and disinfect/sterilize reusable equipment
- Store clean reusable equipment in a sanitary way

- Clean and disinfect the footbath after every client